



SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba•hospital@sai•org•in

NO.SSS/WS/SSBH/STORES/(P/S)/ /2018
DATE:

To,

Sub : - Quotation for Purchase of LED Portable Shadowless Lamp / Spot Light for Shri Saibaba Hospital, Shirdi.

Dear Sir,

We have to purchase **LED Portable Shadowless Lamp / Spot Light** for Shri Saibaba Hospital, Shirdi given in enclosed Annexure "A" & the following terms and Conditions.

Terms and Condition

1. The **LED Portable Shadowless Lamp / Spot Light** will have to be supplied F.O.R. Shirdi.
2. The rate should be inclusive of GST, transport charges, T.O.T. Etc.
3. You will have to mention name of the company, packing, rate per unit, total cost etc.
4. Packing, forwarding, and other charges will not be paid by the Sansthan.
5. The delivery of the ordered material will have to made at Shri Saibaba Hospital, Shirdi on any working days between 10 a.m. to 6 p.m.
6. In case if supply is not made in the proper time and if material is to be purchased from other party at higher rates than the rates given in quotation you will be responsible to compensate the excess amount paid by the Sansthan.
7. You will have to supply the ordered material at the quoted rates in the quotation and no increase will be given or allowed in the rates for any reason whatsoever.
8. In case of breakage / leakage in transit, prior to delivery to Sansthan, you will have to replace the same immediately failing which payment will be made by deducting the cost of the same.
9. Payment will be made by the Sansthan on confirmation of the name of the company, quality, quantity, grade if any etc. as specified in the quotation. No advance payment will be made against the supply.
10. All the Equipments shall be new and manufactured from virgin materials. All the requirements of this supply shall be sourced from the original equipment manufacturer of the model quoted. In case the machine is imported one to import substitution is permitted neither before the wards nor after the award for any part or accessory.
11. Any accessory / accessories not mentioned specifically but required to make system working will have to be supplied without any extra cost.
12. The supplier has to mention the one time visit charges for breakdown and servicing.
13. Equipment mentioned in the Technical specification shall operate on 230V \pm 10%, 50Hz, single phase electric supply if not mentioned specifically.

Page No. 2/-





SHRI SAIBABA SANSTHAN TRUST, SHIRDI

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
Email : saibaba•hospital@sai•org•in

Page No. 2

14. The necessary protective relaying / circuitry shall be there with the machines.
15. The equipment shall be having warranty of 01 year. The warranty and CMC shall cover the batteries and lamps irrespective of whether those are treated as consumables or otherwise.
16. Minimum 4 Nos. of preventive maintenance visits and unlimited breakdown call visits as per requirement must be attended in the warranty and AMC/CMC period. The services to the equipment in warranty and CMC period will be considered as same.
17. Repairs and maintenance training to two Biomedical Engineers should be given for one week or more as per requirement.
18. The equipment must be tropicalized as below-
Working Temperature : Max.45°
Storage Temperature : Max.50°C,
Relative humidity for working : 40 - 60%,
Relative humidity for Storage : Max 90%
19. Among the other things, the responsiveness in Technical Bid will be based on successful demonstration of the offered model of the equipments to SSST officials as mentioned elsewhere in the tender specifications.
20. The supplier has to submit user list with contact telephone numbers for the model quoted.
21. The supplier has to give past performance certificate of the quoted equipment.
22. The supplier should submit all the Datasheets of the quoted model for technical comparison.
23. The product should be time tested, branded and should be OEM.
24. The life of the equipment should be mentioned.
25. The quotation must be clearly mention on the sealed envelope "**LED Portable Shadowless Lamp / Spot Light for Shri Saibaba Hospital, Shirdi**".
26. The quotation in sealed cover may please be submitted to our Shirdi office on or before **Dt. 05/10/2018.**

Thanking you and with blessings of Shri Saibaba.

Yours Sincerely,

Administrative Officer,
Shri Saibaba Sansthan Trust, Shirdi



**SHRI SAIBABA SANSTHAN TRUST, SHIRDI**

SHIRDI: Post: Shirdi - 423109, Tal. Rahata, Dist. Ahmednagar, Maharashtra (India).
 Telephone: +91 - 2423 - 258 671-672-673, Fax: +91 - 2423 - 258660, Telegram: Sainath, Shirdi.
 Email : saibaba•hospital@sai•org•in

Shri Saibaba Hospital, Shirdi
Annexure "A"

Technical Specification for LED Portable Shadowless Lamp / Spot Light
(Preferred Make: - Stann Biomedical, Shree Hospital or equivalent)

Sr. No.	Description	Qty.	Rate Per Inclu. of All Tax	Total Amt. in Rs.																					
1	<p align="center">LED Portable Shadowless Lamp /Spot Light</p> <p>Portable Shadow less Lamp /Spot light having castors and proper balancing with following specifications:</p> <ul style="list-style-type: none"> ❖ It should have cool white intense light ❖ The integrated two step interference filter to remove 99% of the infra red radiation from the light beam ❖ Minimum (<2°C) temperature increase below light head ❖ No build up of heat at the Operator's head ❖ Sealed housing ❖ Environmentally friendly disinfectant resistance powder paint finish. ❖ Integrated ON / OFF switch ❖ Long life LEDs should be used min. 30000 Hrs life ❖ Completely meets the laminar flow purification in the operating room ❖ Excellent shadow less effect and focusing depth ❖ High efficiency endothermal system to control the temperature rising of the operating area. ❖ Large action radius and rotation ranges, all types of movements should be possible as per the user requirement ❖ Changeable Handles should be provided-2 nos. ❖ Utmost stability of position ❖ Input should be protected against high voltage <ul style="list-style-type: none"> • Warranty- 1 Year <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>S. N.</th> <th>Features</th> <th>Requirement</th> </tr> </thead> <tbody> <tr> <td align="center">1</td> <td>Illuminance at 1 mtr. Distance</td> <td>70,000 Lux or more</td> </tr> <tr> <td align="center">2</td> <td>Light field diameter</td> <td>20 – 32 CM</td> </tr> <tr> <td align="center">3</td> <td>Colour temperature</td> <td>4,300 Kelvin</td> </tr> <tr> <td align="center">4</td> <td>Colour rendering index Ra</td> <td>>92</td> </tr> <tr> <td align="center">5</td> <td>Luminance efficiency</td> <td>270 lm/W</td> </tr> <tr> <td align="center">6</td> <td>Reflector</td> <td>Multi facilitated Aluminum or better quality</td> </tr> </tbody> </table> <ul style="list-style-type: none"> • All Technology should be mandatorily with safety approval - CE / UL. Approval certificate is to be submitted for the total unit. • The product should be time tested, branded and should be OEM • AMC Cost- AMC cost should be including Taxes. It should be fixed for next 5 years. • Cost of Consumables to be fixed for warranty period. 	S. N.	Features	Requirement	1	Illuminance at 1 mtr. Distance	70,000 Lux or more	2	Light field diameter	20 – 32 CM	3	Colour temperature	4,300 Kelvin	4	Colour rendering index Ra	>92	5	Luminance efficiency	270 lm/W	6	Reflector	Multi facilitated Aluminum or better quality	1		
S. N.	Features	Requirement																							
1	Illuminance at 1 mtr. Distance	70,000 Lux or more																							
2	Light field diameter	20 – 32 CM																							
3	Colour temperature	4,300 Kelvin																							
4	Colour rendering index Ra	>92																							
5	Luminance efficiency	270 lm/W																							
6	Reflector	Multi facilitated Aluminum or better quality																							

Medical Director
 Shri Saibaba Hospitals, Shirdi

